

# VFC2000

VFC Temperature Data Logger



## PRODUCT USER GUIDE

To view the full MadgeTech product line, visit our website at [madgetech.com](http://madgetech.com).

### TABLE OF CONTENTS

2 Product Overview

2 Device Operation

3 Device Maintenance

3 Need Help?

4 Specifications

4 Ordering Information



# PRODUCT USER GUIDE

## Product Overview

The VFC2000 is a simple solution for vaccine temperature monitoring compliance. Designed to satisfy all CDC and VFC requirements, the VFC2000 provides highly precise, continuous temperature validation. Featuring a convenient LCD screen, the VFC2000 displays current readings, minimum and maximum statistics as well as a battery level indicator. User-programmable alarms trigger an audible and visual alert. Optional AC power source allows battery to be a back up if there is a power loss.



### VFC Requirements

- Detachable, buffered temperature probe
- Out-of-range audible and visual alarms
- Low battery indicator with external power and battery backup
- Current, minimum, and maximum temperature display
- Accuracy of  $\pm 0.5^{\circ}\text{C}$  ( $\pm 1.0^{\circ}\text{F}$ )
- Programmable logging interval (1 reading per second to 1 reading per day)
- Daily check reminder warning
- Suitable for vaccine transportation
- Also monitors ambient room temperature

## Device Operation

1. Install the MadgeTech 4 Software onto a Windows PC.
2. Connect the data logger to the Windows PC with the provided USB cable.
3. Launch the MadgeTech 4 Software. The VFC2000 will appear in the Connected Devices window indicating the device has been recognized.
4. Select the start method, reading rate and any other parameters appropriate for the desired data logging application. Once configured, deploy the data logger by clicking the Start icon.
5. To download data, select the device in the list, click the Stop icon, and then click the Download icon. A graph will automatically display the data.

### Selection Buttons

The VFC2000 is designed with three selection buttons:



**Scroll:** Allows user to scroll through current readings, average stats, daily minimum and maximum temperatures and device status information displayed on the LCD screen.



**Units:** Allows users to change displayed units of measurement to either Celsius or Fahrenheit.



**Start/Stop:** To activate Manual Start, arm the device through the MadgeTech 4 Software. Hold the button for 3 seconds. There will be two beeps confirming the device is started. Reading will be displayed on the screen and status in software will change from **Waiting to Start** to **Running**. To pause logging while running, hold button for 3 seconds.

# PRODUCT USER GUIDE

## LED Indicators



**Status:** Green LED blinks every 5 seconds to indicate the device is logging.



**Check:** Blue LED blinks every 30 seconds to indicate daily statistic check has gone past 24 hours. Hold Scroll button for 3 seconds to reset reminder.



**Alarm:** Red LED blinks every 1 second to indicate an alarm condition is set.

## Device Maintenance

### Battery Replacement

**Materials:** U9VL-J Battery or any 9 V Battery

1. On the bottom of the data logger, open the battery compartment by pulling in on the cover tab.
2. Remove the battery by pulling it from the compartment.
3. Install the new battery, taking note of the polarity.
4. Push the cover closed until it clicks.

### Recalibration

Recalibration is recommended annually or bi-annually for any data logger; a reminder is automatically displayed in the software when the device is due. To send devices back for calibration, visit [madgetech.com](http://madgetech.com).



## Product Support & Troubleshooting:

- Refer to the Troubleshooting section of this document.
- Visit our Knowledge Base online at [madgetech.com/resources](http://madgetech.com/resources).
- Contact our friendly Customer Support Team at (603) 456-2011 or [support@madgetech.com](mailto:support@madgetech.com).



## MadgeTech 4 Software Support:

- Refer to the built-in help section of the MadgeTech 4 Software.
- Download the MadgeTech 4 Software Manual at [madgetech.com](http://madgetech.com).
- Contact our friendly Customer Support Team at (603) 456-2011 or [support@madgetech.com](mailto:support@madgetech.com).

# PRODUCT USER GUIDE

## Specifications

Specifications are subject to change without notice. Specific warranty remedy limitations apply. Call (603) 456-2011 or go to [madgetech.com](http://madgetech.com) for details.

TEMPERATURE	
<b>Temperature Range</b>	-20 °C to +60 °C (-4 °F to +140 °F)
<b>Resolution</b>	0.01 °C (0.018 °F)
<b>Calibrated Accuracy</b>	±0.50 °C/± 0.18 °F (0 °C to +55 °C/32 °F to 131 °F)
<b>Response Time</b>	10 minutes free air

REMOTE CHANNEL	
<b>Thermocouple Connection</b>	Female subminiature (SMP) (MP model) Pluggable screw terminal (TB model)
<b>Cold Junction Compensation</b>	Automatic, based on internal channel
<b>Max. Thermocouple Resistance</b>	100 Ω
<b>Thermocouple K</b>	<b>Range (°C):</b> -65 °C to +40 °C <b>Resolution:</b> 0.1 °C <b>Accuracy:</b> ±0.5 °C
<b>Response Time</b>	τ = 2 minutes to 63% of change

GENERAL	
<b>Reading Rate</b>	1 reading every second up to 1 reading every 24 hours
<b>Memory</b>	16,128 readings
<b>LED Functionality</b>	3 status LEDs
<b>Wrap Around</b>	Yes
<b>Start Modes</b>	Immediate and delay start
<b>Calibration</b>	Digital calibration through software
<b>Calibration Date</b>	Automatically recorded within device

<b>Battery Type</b>	9 V lithium battery included; user replaceable with any 9 V battery
<b>Battery Life</b>	3 years typical at 1 minute reading rate
<b>Data Format</b>	<b>For Display:</b> °C or °F <b>For Software:</b> Date and time stamped °C, K, °F or °R
<b>Time Accuracy</b>	± 1 minute/month
<b>Computer Interface</b>	USB to mini USB, 250,000 baud for standalone operation
<b>Operating System Compatibility</b>	Windows XP SP3 or later
<b>Software Compatibility</b>	Standard Software version 4.2.21.0 or later
<b>Operating Environment</b>	-20 °C to +60 °C (-4 °F to +140 °F), 0 %RH to 95 %RH non-condensing
<b>Dimensions</b>	3.0 in x 3.5 in x 0.95 in (76.2 mm x 88.9 mm x 24.1 mm) Data logger only
<b>Glycol Bottle</b>	30 mL
<b>Probe Length</b>	72 in
<b>Material</b>	ABS Plastic
<b>Weight</b>	4.5 oz (129 g)
<b>Approvals</b>	CE
<b>Alarm</b>	User configurable high and low audible and on-screen alarms. Alarm Delay: A cumulative alarm delay may be set in which the device will activate the alarm (via LED) only when the device has recorded a user specified time duration of data.
<b>Audible Alarm Functionality</b>	1 BEEP per second for reading alarm above/ below threshold

**BATTERY WARNING:** BATTERY MAY LEAK, FLAME OR EXPLODE IF DISASSEMBLED, SHORTED, CHARGED, CONNECTED TOGETHER, MIXED WITH USED OR OTHER BATTERIES, EXPOSED TO FIRE OR HIGH TEMPERATURE. DISCARD USED BATTERY PROMPTLY. KEEP OUT OF REACH OF CHILDREN.

## Ordering Information

<b>VFC2000</b>	PN 902238-00	VFC Temperature data logger with USB to mini USB cable
<b>Power Adapter</b>	PN 901839-00	Replacement USB universal power adapter
<b>U9VL-J</b>	PN 901804-00	Replacement battery for VFC2000



6 Warner Road, Warner, NH 03278  
(603) 456-2011  
info@madgetech.com  
madgetech.com